



सॉफ्टवेयर टेक्नोलॉजी पार्क्स ऑफ इंडिया  
**SOFTWARE TECHNOLOGY PARKS OF INDIA**  
(भारत सरकार के इलेक्ट्रॉनिकी और सूचना प्रौद्योगिकी मंत्रालय के अन्तर्गत एक  
स्वायत्त सोसाइटी, इलेक्ट्रानिक सिटी, बेंगलूरु – 100)  
**(An Autonomous Society under Ministry of Electronics  
and Information Technology**  
**Govt. of India, Electronic City, Bengaluru -100)**

## शुद्धि-पत्र / CORRIGENDUM

This has reference to the Tender Ref:- STPIB/ADMIN/PSG/2020-21/02 Dated: 23.07.2020 published for Supply, Installation, Testing & Commissioning (SITC) of Healthcare Equipment for AIC STPI Bengaluru project under AIC STPINEXT INITIATIVES with comprehensive warranty for 3 years.

The Corrigendum has been issued for the following:

### A. 3D Printer:

Clause No	Basic Description	Machine Configuration	To be read as
Part I (e) Appendix - A Technical Specification Serial No. 03 3D printer Table no 7.	Operation and Process Control	Machine should build parts in Baffled oven type heated chamber with Uniform Temperature control. The oven should have two or more independent air heaters to create a uniform air heated build environment throughout the volume of the build with temperature capacity of 220C or better.	Machine should build parts in Baffled oven type heated chamber with Uniform Temperature control. The oven should have two or more independent air heaters to create a uniform air heated build environment throughout the volume of the build with temperature capacity ranging from 200C to 250C or better.
Part I (e) Appendix - A Technical Specification Serial No. 03 3D printer Table no 9 - 9.1 .	Operational and Facility Requirements	Noise level of the machine at any point (idle/operational) should be less than 46dbA. Relevant documentation/test results to be provided.	Noise level of the machine at any point (idle/operational) should be less than 50 dbA. Relevant documentation/test results to be provided.

**B. General :**

<b>Clause No</b>	<b>Description of the RFT</b>	<b>To be read as</b>
1.1 Part -1 Technical Bid document Table SL.No 5	Bidder should have the valid ISO 9001: 2015, 14001:2015 & 18001:2007 Certificates	Relaxation of the clause for submitting ISO certificates. (This clause is dropped).
Part I(a) - General Terms & Conditions 12.2 (b)	In case, any arithmetical /numerical /typo error is found in the commercial quoted by the bidder in the BoQ such as not quoting the GST price, mismatch in the unit price or total price will not be entertained during the bidding. The lowest price reflected in the BoQ column No. 13 shall be considered for evaluation.	In case, any arithmetical /numerical /typo error is found in the commercial quoted by the bidder in the BoQ such as not quoting the GST price, mismatch in the unit price or total price will not be entertained during the bidding. The lowest price reflected in the BoQ column No. 14 shall be considered for evaluation.

**This is for your kind information.**

**निदेशक/ Director**  
**एसटीपीआई-बेंगलूरु/STPI - Bengaluru**

**दिनांक/Date : 17.08.2020**